

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Converse Laboratories, Inc.  
**Project:** Southside Water Users  
**Sample Matrix:** Drinking Water

**Service Request:** R2602093  
**Date Collected:** 02/16/26 12:55  
**Date Received:** 02/17/26 18:19

**Sample Name:** Ives St.  
**Lab Code:** R2602093-002

**Units:** ug/L  
**Basis:** NA

Trihalomethanes, (GC/MS)

**Analysis Method:** 524.2

Analyte Name	Result	MRL	Dil.	Date Analyzed	Q
Bromodichloromethane	1.9	0.50	1	02/18/26 19:40	
Bromoform	0.50 U	0.50	1	02/18/26 19:40	
Chloroform	13	0.50	1	02/18/26 19:40	
Dibromochloromethane	0.50 U	0.50	1	02/18/26 19:40	
Trihalomethanes, Total	15	2.0	1	02/18/26 19:40	

Surrogate Name	% Rec	Control Limits	Date Analyzed	Q
1,2-Dichlorobenzene-d4	105	70 - 130	02/18/26 19:40	
Bromofluorobenzene	78	70 - 130	02/18/26 19:40	

## Stage 2 TTHM Summary Sheet Location Running Annual Aveage

PWS Name: Sothside Water Inc.  
 Contact Person: James V Lettiere Jr.  
 Contact person Phone#: cell: (315) 486-7937

PWS Numbr: NY222133  
 Date Sampled 02/16/26  
 Date Analyzed 03/03/26

Sample Site	TTHM Individual Results (ug/L)				Total TThm (ug/L)
	Chloroform	Bromoform	Bromodichloromethane	Dibromodichloromethan	
Last Hydrant on Ives St.	13		1.9		14.9

Sample Site: MRT01		
Quarter	Month/year	Total TTHM
Present	Jan. - March. 2025	14.9
Previous	Oct- Dec. 2025	42.8
Previous	July- Sept. 2025	82.1
Previous	April - June. 2025	57.6
TTHM Locational Running Average:		49.4

Sample Site:		
Quarter	Month/year	Total TTHM
Present		
Previous		
Previous		
Previous		
TTHM Locational Running Average:		

Sample Site:		
Quarter	Month/year	Total TTHM
Present		
Previous		
Previous		
Previous		
TTHM Locational Running Average:		

Sample Site:		
Quarter	Month/year	Total TTHM
Present		
Previous		
Previous		
Previous		
TTHM Locational Running Average:		

Name of Laboratory: Converse Laboratories Inc./AES  
 Address: 800 Starbuck Ave  
Watertown Ny 13601  
Phone 315-788-8388

**Notes: At this Site Chlorine Residual was .30 when sampled**  
 Locatin running average (LRAA)= Summation of the site specific sample results for the current quarter and previous three (3) quarters, divided by four (4)

LRAA=           (Present qtr.+ Previous Qtr. + Previous Qtr. + Previous Qtr.)

**ALS Group USA, Corp.**  
dba ALS Environmental

Analytical Report

**Client:** Converse Laboratories, Inc.  
**Project:** Southside Water Users  
**Sample Matrix:** Drinking Water

**Service Request:** R2602093  
**Date Collected:** 02/16/26 12:40  
**Date Received:** 02/17/26 18:19

**Sample Name:** Holcomb St.  
**Lab Code:** R2602093-001

**Units:** ug/L  
**Basis:** NA

**Haloacetic Acids (GC)**

**Analysis Method:** 552.2  
**Prep Method:** Method

Analyte Name	Result	MRL	Dil.	Date Analyzed	Date Extracted	Q
Bromoacetic Acid	1.0 U	1.0	1	03/03/26 22:50	3/2/26	
Chloroacetic Acid	2.0 U	2.0	1	03/03/26 22:50	3/2/26	
Dichloroacetic Acid	<b>5.3</b>	1.0	1	03/03/26 22:50	3/2/26	
Trichloroacetic Acid	<b>8.0</b>	1.0	1	03/03/26 22:50	3/2/26	
Dibromoacetic Acid	1.0 U	1.0	1	03/03/26 22:50	3/2/26	
Haloacetic Acids, Total	<b>13</b>	6.0	1	03/03/26 22:50	3/2/26	

Surrogate Name	% Rec	Control Limits	Date Analyzed	Q
2,3-Dibromopropionic Acid	86	70 - 130	03/03/26 22:50	

**Stage 2 HAA5 Summary Sheet  
Location Running Annual Aveage**

PWS Name: Sothside Water Inc.  
 Contact Person: James V Lettiere Jr.  
 Contact person Phone#: cell: (315) 486-7937

PWS Numbr: NY222133  
 Date Sampled 02/16/26  
 Date Analyzed 02/18/26

Sample Site	HAA5 Individual Results (ug/L)						Total HAA5 (ug/L)
	Chloacetic Acid	Dichloacetic acid		Bromoacetic acid	Trichloroacetic acid	Dibromoacetic acid	
Last House on Holcomb St. (EP)	U	5.3		U	8	U	13.3

Sample Site:EP01		
Quarter	Month/year	Total HAA5
Present	Jan. - March. 2026	13.3
Previous	Oct- Dec. 2025	19.7
Previous	July- Sept. 2025	40.4
Previous	April - June. 2025	35.7
HAA5 Locational Running Average:		27.3

Sample Site:		
Quarter	Month/year	Total HAA5
Present		
Previous		
Previous		
Previous		
HAA5 Locational Running Average:		

Sample Site:		
Quarter	Month/year	Total HAA5
Present		
Previous		
Previous		
Previous		
HAA5 Locational Running Average:		

Sample Site:		
Quarter	Month/year	Total HAA5
Present		
Previous		
Previous		
Previous		
HAA5 Locational Running Average:		

Name of Laboratory: Converse Laboratories Inc./AES  
 Address: 800 Starbuck Ave  
Watertown Ny 13601  
Phone 315-788-8388

Notes:  
 Locatin running average (LRAA)= Summation of the site specific sample results for the current quarter and previous three (3) quarters, divided by four (4)

LRAA= (Present qtr.+ Previous Qtr. + Previous Qtr. + Previous Qtr.)